🗟 | HEALTH

Heat Shrink Tubing

Custom and Standard Polyolefin Heat Shrink Solutions

Product & Process Consistency for Optimal Clinical Outcome

Heat shrink tubing applications range from device bonding – including balloon bonding, lab bonding, butt bonding – to protective sheathing, marker banding and strain reliefs.

Lubrizol Life Science Health's (LLS Health) precision heat shrink tubing capabilities yield cost-effective solutions for a broad range of medical devices. Our high shrink ratios ensure a high compression force at low recovery temperatures (275 - 350°F/135 - 177°C). Within an ISO 13485: 2016 certified facility, our process and process control of polyolefin heat shrink enables excellent concentricity and flexible thin-wall construction facilitating easy of use, excellent fit and compliance.

Readily Available Stock Sizes & Design Specific Customization

Whether you require a range of standard heat shrink tubing to meet project timelines or custom sizing to achieve precision fits, LLS Health offers the best solution. Because heat shrink tubing recovers to a smaller diameter after heating, we put the proper process controls in place to ensure that every product meets your requirements for minimum expanded inner diameter, maximum recovered inner diameter, recovered wall thickness and length.

Plus, we offer a full range of polyolefin materials, including LDPE, MDPE and LLDPE (low-, mediumand linear low-density polyethylene) options to meet your specific design and process needs.

Exceptional Manufacturing Capabilities

LLS Health's experienced staff utilizes high concentricity extrusion, precise radiation and a tightly controlled expansion process to tackle the most challenging heat shrink tubing requirements. Our expertise includes:

- Standard and custom sizes
- Expanded ID 0.01"-0.30" (0.25 mm 7.62 mm)
- Expanded wall thicknesses as thin as 0.0003" (0.00762 mm)
- Length: 0.5"- 60" (1.2 cm 152.4 cm), continuous length for some sizes
- Shrink ratios as high as 15:1
- Rapid and low recovery temperatures of 275 350°F/ 135 – 177°C, depending on material
- Colors: clear and customer-specified
- Printing



LLS Health's Standard Polyolefin Heat Shrink Tubing Line

View Our Product Line

Part Number	10032-CS4L	10056-CS4L	10080-CS4L	10092-CS4L	10120-CS4L	10160-CS4L	10280-CS4L	10490-CS7L
Minimum	0.032"	0.056"	0.080"	0.092"	0.120"	0.160"	0.280"	0.490"
Expanded ID	(0.81 mm)	(1.42 mm)	(2.03 mm)	(2.34 mm)	(3.05 mm)	(4.06 mm)	(7.11 mm)	(12.45 mm)
Maximum	0.013"	0.019"	0.025"	0.028"	0.036"	0.045"	0.077"	0.094"
Recovered ID	(0.33 mm)	(0.48 mm)	(0.64 mm)	(0.71 mm)	(0.91 mm)	(1.14 mm)	(1.96 mm)	(2.39 mm)
Expanded	0.00083"	0.00094"	0.00085"	0.00085"	0.00093"	0.00094"	0.001"	0.00085"
Wall	(0.021 mm)	(0.024 mm)	(0.022 mm)	(0.022 mm)	(0.024 mm)	(0.024 mm)	(0.025 mm)	(0.021 mm)
Maximum	0.0028"	0.0028"	0.004"	0.00325"	0.003"	0.0039"	0.008"	0.008"
Recovered Wall	(0.071 mm)	(0.071 mm)	(0.102 mm)	(0.083 mm)	(0.076 mm)	(0.099 mm)	(0.203 mm)	(0.203 mm)

Determine the Expanded ID

Select a product with an expanded ID large enough to easily fit over the largest OD of your component.



About Lubrizol Life Science Health

The Health business team partners with customer to speed their innovative medical devices and differentiated pharmaceutical products to market. Our dedicated team provides best-in-class polymers and excipients, along with state-of-the-art product design, development, and manufacturing services, with the ultimate goal of creating solutions that improve patient outcomes.

For more information visit www.Lubrizolmeddevicecm.com or email lls_cm@lubrizol.com.



9911 Brecksville Road Cleveland, OH 44141-3201 USA

Email: LLS_CM@lubrizol.com | Web: Lubrizolmeddevicecm.com

The information contained herein is believed to be reliable, but no representations, guarantees or warranties of any kind are made as to its accuracy, suitability for particular applications or the results to be obtained. The information often is based on laboratory work with small-scale equipment and does not necessarily indicate end-product performance or reproducibility. Formulations presented may not have been tested for stability and should be used only as a suggested starting point. Because of the variations in methods, conditions and equipment used commercially in processing these materials, no warranties or guarantees are made as to the suitability of the products for the applications disclosed. Full-scale testing and end-product performance are the responsibility of the user. Lubrizol Advanced Materials, Inc.'s direct control. The SELLER MAKES NO WARRANTIES, EXPRESS OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. Nothing contained herein is to be considered as permission, recommendation nor as an inducement to practice any patented invention without permission of the patent owner. Lubrizol Advanced Materials, Inc., is a wholly owned subsidiary of The Lubrizol Corporation.

@2020 The Lubrizol Corporation, all rights reserved. All marks are the property of The Lubrizol Corporation. The Lubrizol Corporation is a Berkshire Hathaway company.